Leader in Biomolecular Solutions for Life Science

www.abclonal.com

Rabbit anti-Human IL-2 mAb

Catalog No.: RMK0516

Basic Information

Catalog No.
RMK0516

Catagory
Elisa Antibody Kit

Application
ELISA,CLIA

Product Information

Ig Type
Rabbit IgG

Purification
Affinity purification

Endotoxin Level
N/A

Storage
Store at -20℃
Avoid repeated freeze-thaw cycles.

Formulation
Supplied as a 0.2 μm filtered solution in PBS with 0.05% Proclin 300, pH 7.3.

Contact

order@abclonal.com

support@abclonal.com

www.abclonal.com

Background

This gene is a member of the interleukin 2 (IL2) cytokine subfamily which includes IL4, IL7, IL9, IL15, IL21, erythropoietin, and thrombopoietin. The protein encoded by this gene is a secreted cytokine produced by activated CD4+ and CD8+ T lymphocytes, that is important for the proliferation of T and B lymphocytes. The receptor of this cytokine (IL2R) is a heterotrimeric protein complex whose gamma chain is also shared by IL4 and IL7. The expression of this gene in mature thymocytes is monoallelic, which represents an unusual regulatory mode for controlling the precise expression of a single gene. The targeted disruption of a similar gene in mice leads to ulcerative colitis-like disease, which suggests an essential role of this gene in the immune response to antigenic stimuli.

Immunogen Information

Immunogen
Recombinant Human IL-2 protein

Cross-Reactivity


Assay Applications

Human IL-2 Sandwich Immunoassay


		Recommended Concentration	Sample
ELISA	Capture	1-4 μg/ml	Rabbit anti-Human IL-2 mAb (Cat. No. RMK0516)
	Detection	0.05-0.2 μg/ml	Rabbit anti-Human IL-2 mAb (Cat. No. RMK0517)
	Standard	1.56-100 pg/mL	Recombinant Human IL-2 Protein (Cat. No. RP01039)

CLIA	Capture	0.5-2 μg/ml	Rabbit anti-Human IL-2 mAb (Cat. No. RMK0516)
	Detection	0.125-0.5 μg/ml	Rabbit anti-Human IL-2 mAb (Cat. No. RMK0517)
	Standard	4.9-5000 pg/mL	Recombinant Human IL-2 Protein (Cat. No. RP01039)


Validation Data



No significant cross-reactivity or interference was observed.



Placed at 37°C for 7 days, the stability of the standard curve all conform to CV <10%.



This standard curve is only for demonstration purposes. A standard curve should be generated for each assay.